



MICROTESTMACHINES Co.

Langmuir–Blodgett trough

LT-103

Operating Manual

2014



Attention! In view of constant improvements of the device and updating of its mechanical and electronic parts some differences may exist between real details and the illustrations, that, however, does not mean deterioration of quality and characteristics of the complex.



Microtestmachines Co.,
Telman St. 44 of. 6, Gomel, 246050, Belarus
Tel./Fax: +375 232 715463
E-mail: microtm@mail.ru

CONTENTS

1	GENERAL DESCRIPTION.....	4
1.1	LB trough	4
1.2	Surface pressure sensor	5
1.3	Dipper unit.....	7
1.4	Control electronic unit	8
2	LT-103 CONTROL SOFTWARE	11
2.1	Main window	11
2.2	Panel of Experimental Setup.....	13
2.3	Control panels	15
2.4	Graph plotting window	17
2.5	Configuration panels	18
3	OPERATIONS WITH CONTROL ELECTRONIC UNIT	20
3.1	Sensor adjustment.....	20
3.2	Sensor calibration	21
3.3	Manual control of motors.....	23
4	PRACTICAL GUIDE TO OPERATION WITH EXPERIMENTAL DEVICE FOR FILM APPLICATION	
	LT-103	24
4.1	Switching the device on and calibration.....	24
4.2	Measurement of compression isotherms.....	26
4.3	Deposition of mono- and multimolecular films.....	27
4.4	Data processing.....	30

1 GENERAL DESCRIPTION

Device for film application LT-103 is intended for mono- and multimolecular film deposition on surface of solid Substrates according to Langmuir–Blodgett (LB) technique.




Fig. 1.1 – General view of the Langmuir–Blodgett trough LT-103. Left – LB trough; right – control electronic unit.

Langmuir–Blodgett trough LT-103 (Fig. 1.1) consists of following units

1. LB trough;
2. Dipper unit;
3. Surface pressure sensor;
4. Control electronic unit with set of connection cables.

Additional accessories include:

1. a PTFE holder for substrate clamping;
2. a PTFE tripod sample table for placement in dipping well of the trough to deposit film by liquid draining technique;
3. PTFE blocks for inserting into the trough to decrease liquid volume and placement of special dopes;

For the device operation, it is necessary to allocate additionally host PC to run the device control software. The host PC should operate under Win32 operating system and contain a free serial port (COM port) to connect with the device control electronic unit.

1.1 LB trough

Schematic diagram of the LB trough is shown in Fig. 1.2. LB trough itself is made of PTFE block. Free surface area of the trough is 400 cm² (290x138 mm) and depth is 30 mm. Dipping well installed at left side of the trough has diam. 60 mm and own depth 45 mm that allows to apply films to the samples as long as 60 mm. The trough holds about 1 l of liquid. Barrier for constraining free surface

of liquid is made as a PTFE bar and slides on ledge in the trough top. Taking into account dead area behind the barrier, maximum effective area of liquid free surface in working zone is about 358 cm².




Fig. 1.2. Top view schematic diagram of Langmuir-Blodgett trough LT-103.

From bottom side, the trough is equipped with a jacket for liquid thermostating by external device (thermostat not included). For monitoring temperature inside the trough, a thermal sensitive element is embedded in rear wall of the PTFE trough near the surface.

The PTFE trough is mounted in a frame of aluminum alloy. From rear side of the frame, a mechanism of the barrier horizontal movement is installed. Drive shaft is powered by a stepper motor that provides precise horizontal positioning of the barrier.

Two vertical stands at the left of the frame are used for installation of surface pressure sensor and dipper unit. All the installed units allow their turning around the vertical stands and vertical movement along the axes. To fix the units in desirable position on the stands, they are equipped with special screws.

Hose for draining the liquid off the trough is attached to nipple from bottom side of dipping well. Jacket for thermostating is also equipped from right side with two hose nipples for heat carrying liquid input and output.

1.2 Surface pressure sensor

Surface pressure sensor serves to monitor surface pressure of liquid in the trough. This information is then used to form error signal for feedback system of the device.

The sensor unit is mounted on the stand at left side of the PTFE trough frame according to scheme in Fig. 1.2. Schematic diagram of the sensor is shown in Fig 1.3. Sensitive element of the sensor is made as a flexible cantilever (leaf spring of steel). To the free end of the cantilever, a Wilghelmy plate is hung to be put into liquid and to serve as initial sensor. The plate is a piece of filter paper featuring good wetting with the liquid (not shown). Size of the plate should be about 14x14 mm to provide area about 200 mm² (this value is considered as standard for the device measuring system).




Fig. 1.3 – Schematic diagram of surface pressure sensor.

Bend of the cantilever depends on force S exerted by liquid surface to Wilghelmy plate. Change of surface pressure (tension) causes change of force S and, correspondingly, deflection of the cantilever free end off the neutral position along vertical axis.

To detect sensitive cantilever deflection, a laser-beam scheme is applied (Fig. 1.3). It comprises laser source and two-section photodetector. Laser beam from the source is targeted onto mirror mounted on rear side of the cantilever free end. After reflection off the mirror, the beam gets to two-sectioned photodetector. Position of reflected spot on the photodetector sensitive area depends on cantilever vertical deflection. Electronic system compares output signals from upper and lower segments of the photodetector and determines magnitude of the cantilever deflection that in its turn describes changes of surface pressure.

Design of the surface pressure sensor provides for smooth adjustment of the cantilever spring constant by change of free end length using a sliding slat (Fig. 1.3). Installation of flexible cantilever with different thickness, width or material gives an opportunity of step-like change of the spring constant. Combination of above two ways provides possibility to select and adjust necessary range of the sensor unit sensitivity.

Laser source of the sensor is mounted on spring-loaded platform. Laser module includes power supply circuit, laser diode and set of lenses in a tube case. The laser module has following characteristics:

aperture	3 mm
focus length	40 mm
wavelength	650 nm $\pm 5\%$
laser diode power output	< 5 mW
power supply	DC 5 V

Position of the platform with laser module can be adjusted to target the beam to mirror on rear side of the cantilever. Fine adjustment of the platform position to target to the mirror and its placement in the plane perpendicular to photodetector can be performed by rotation of three adjustment screws and turning the platform around its pivot point.

Photodetector is installed in specialized mechanism providing the photodiode sensitive area positioning right opposite laser beam reflected off the mirror on the cantilever rear side. To adjust photodetector position operator should rotate adjustment screw in the hole on top of the sensor unit. Inside the sensor unit a preamplifier is installed near the photodetector to reduce interference and noise in the output signal. Preamplifier provides initial amplification and processing of the photodetector output signal.

1.3 Dipper unit

Dipper unit is a specialized mechanism for lifting and dipping the substrates into liquid. The unit is mounted on the stand at left side of the PTFE trough frame. Schematic diagram of the unit is shown in Fig. 1.4. Dipper unit provides vertical stroke of the hanging arm 85 ± 2 mm.




Fig. 1.4 – Schematic diagram of dipper unit.

For the sample (substrate) easy fixation on the hang arm of dipper unit use clamping PTFE holder that can be attached to threaded holes in the arm (Fig. 1.5).



Fig. 1.5 – A PTFE holder for substrate clamping.

The dipping/lifting mechanism is driven by step motor controlled by control electronic unit. Connection cable of the dipper unit is to be plugged to socket at left wall of the trough frame under

the stands. If necessary, manual vertical positioning of the hang arm can be done with knob on top of the dipper unit.

Figure 1.6 shows the trough with all installed units necessary for the device operation.




Fig. 1.6 – LB trough filled with liquid and equipped with surface pressure sensor and dipper unit bearing substrate ready for operation. Barrier is in rightmost (initial) position.

Dipper unit has two mounting blocks placed on its opposite sides (front and rear). That allows placement of the unit on its stand with hang arm directed off the trough to perform programmable dipping of any samples in vessels outside the LB trough. Besides, external stand (rod diam. 8 mm) located at the trough (distance is restricted by the dipper connection cable length) can be also used for mounting the dipper unit to perform any suitable tasks of sinking and lifting the sample attached to the hang arm.

1.4 Control electronic unit

Control electronic unit is a specialized unit for control over the device for film application according to the commands from host PC. Front panel of the unit bears LC status display for displaying basic parameters of the device systems and a keypad that allows direct setting of basic functions without control software on host PC (Fig. 1.7). Control electronic unit is connected with LB trough using special cables and with host PC using standard serial port cable. To power the unit, it should be plugged to electrical outlet (grounded 220 V 50 Hz) using power cord. All these cables are supplied. Switching of the unit (and the entire device) should be made by power switch on its rear side.



CAUTION! Socket of electrical outlet must have reliable grounding. If there is no grounding in the socket, you are to provide grounding for the electronic unit and LB trough to prevent electric shock at the device operation.

After switching the control electronic unit on, initial mode of the device status will be shown in the display – **normal mode**. Upper line of the display under this mode shows current readings of the surface pressure sensor. Pressing corresponding buttons on the keypad, all other modes and functions of the device tuning and control can be entered.







Fig 1.7 – Control panel of the device electronic unit.



Buttons of the keypad on control panel have following functions (when it is in initial state – normal mode):


1	enter calibration mode to define minimum value;
2	enter calibration mode to define maximum value;
3	display feedback service information;
4	display current temperature of environment and inside the trough;
5	display steps performed by stepper motors;
9	toggle save mode on/off for the pressure sensor laser (when the mode is on, laser will be automatically switched off after 5 min of the device idle time;
Enter	display output signal from photodetector segments (enter mode of adjusting laser detection system of the surface pressure sensor). Processed signal values in upper line are <Integral value> <Differential value> . Additional pressing of button Enter on the keypad toggles format of the output signal displayed in upper line to <ADC value from segment A> < ADC value from segment B> . That shows distribution of cumulative output signal received from lower A and upper B segments of the photodetector. Lower line shows prompt A photo diode B . Under other modes, button Enter is used for confirmation of entered number


values;


 return to initial mode (go up one menu level);

 start barrier movement to left side (free area confinement). To stop motion, press button  on the keypad or wait till barrier reaches leftmost position;

 start barrier movement to right side (free area increase). To stop motion, press button  on the keypad or wait till barrier reaches rightmost position;

 move dipper unit hang arm downward (motion is in progress while the button is pressed);

 move dipper unit hang arm upward (motion is in progress while the button is pressed);

 show scrolling prompt line in the display with brief description of the keypad button functions;

6–8, 0, . are not in use for device control.

Number keys (1 to 9 and 0) and decimal point button of the keypad on control electronic unit are used for entering numbers under corresponding modes of the device parameters setting.

2 LT-103 CONTROL SOFTWARE

Experimental device for film application LT-103 operates under the control of a specialized software **LB Trough control** delivered in the complete set and run on the IBM PC-compatible computer. The control software is installed on the host PC by executing file `lbt2inst.exe` with the setup procedure. After the installation, program **LB Trough control** can be started by running file `lbt.exe` in folder where the software locates on hard disk or by other ways used in the operating system (e.g. double click on the program shortcut in Windows desktop)

2.1 Main window

After start of **LB Trough control** software, main program window appears in the Windows desktop (Fig. 2.1). Main window contains a menu line at its top, toolbar under the menu line, working area in its middle portion and status line at the bottom. Working area is a place where control panels and other interface elements (graph windows, dialogues) are brought up for operation with the device and measured data.




Fig. 2.1 – Main window of **LB Trough control** program.

Control panels **Deposition Control**, **Isotherm Control**, **Dipping Control** and **Manual**

Control Unit can be docked to right border of the working area (Fig. 2.1) to provide more space for other tasks. To dock a panel, drag it by its title to area at the right border. To undock panel, drag it off the docking area fixing corresponding tab at the right border. Working area can be scrolled with scroll bars appeared when panels or windows get outside existing frame.

Program menu provides access to all the functions of control software. Menu item 'File' contains commands necessary for operations with files, general settings of the program and command to exit program (Fig. 2.2 a). Menu item 'View' contains commands for bringing up specialized control panels for operation with the device. Menu item 'Device' provides commands for operations with the hardware: connection with device, disconnection of the device, device initialization, and settings for the communication port parameters.




Fig. 2.2 – Content of menu items 'File' (a), 'View' (b) and 'Device' (c).

Menu item 'Window' contains standard commands for operations with open windows in the program working area. Menu item 'Help' provides service identification information about the connected device and information about program itself.

Program toolbar contains buttons for quick access to main operations: file save and open dialogues, activation of specialized control panels necessary for work with the device (Fig. 2.3). Toolbar can be detached off the toolbar area under the menu line and placed as floated panel within the desktop.




Fig. 2.3 – Functions of buttons in the program toolbar.

At bottom border of the program main window, a status line is located (Fig. 2.4). It provides service information about:

- current device communication status (**Connected** or **Disconnected**);
- current surface pressure readings (**SP:###**);
- current barrier position (**BPos: ###**);
- current value of free liquid area (**Square: ###**);
- current dipper mechanism position (**DPos: ###**);
- currently executed process or its progress status (**Process: <Description>**).



Fig. 2.4 – Example of information in the program status line.

General **Indicator** panel is always visible in the program working area (Fig. 2.5 left). It repeats information displayed in the program status line:

- current surface pressure readings (SP);
- current barrier position (BPos);
- current value of free liquid area (Square);
- current dipper mechanism position (DPos);
- currently executed process or its progress status (Process).

Indicator panel has no controls for hiding it and also a special command for bringing it up. But it is not always on top and can be overlapped by other control panels in the working area.




Fig. 2.5 – General indicator panel (left) and indicator of deposition progress (right).

Additionally, when deposition is executed specialized indicator graphically and in text shows progress of the procedure (Fig. 2.5 right):

- left column – relative position of dipper hang arm (100% is upper limit, 0 is lower limit);
- middle column – percentage of deposition program fulfilment (currently deposited layer of total number);
- right column – current free surface as a percentage of full working area. Full working area is 100%. As deposition is running, free surface area decreases that indicates decrease of deposited substance amount in the trough.

2.2 Panel of Experimental Setup

Experimental Setup panel (Fig. 2.6) serves as measured data organizer. It is always in the working area but it can be minimized within it. Two tabs of the panel contain means for describing measured data and used composition (tab **Description**) and controls over the data sets (tab **Data**).

Tab **Description** provides:

- general description of the experiment (group –Data List–);
- assignment of working folder on hard disk for data storing (group –Working Folder–). It is recommended to assign (type or browse to select existing) working folder right after first start of the control program;
- description of the deposited substance (group –Solution Composition–). Table in right part of group –Solution Composition– allows to enter information about the deposited composition constituents or single substance if number of constituents (Number of constituents) is one.

Tab **Data** provides:

- list of experiment sets registered in the program and their description (table –Experiment Registry–). Column Loaded of the table shows the experiment status in current session. To work with the experiment sets in program, they should be loaded (see description of controls below). If necessary, shown values can be manually edited right in the table cells;
- switch between experimental data sets (selector –Data set–). Each experimental data set (data measured at one procedure cycle) is treated as a separate set. One file can contain a series of data sets.
- a viewer pane for displaying selected set of experimental data (table –Data sheet–). If necessary, data values can be manually edited right in the cells;

Experimental Setup - Description Tab

Data List

User name: Dr. Carter
 Compound name: Cholesterol
 Substrate: silicon
 Subphase: Diwater
 Temperature: 20

Working Folder: D:\s LBTrough2\data

Solution Composition

Number of constituents: 2
 Volume of Solvent, ml: 0.150
 Aliquot, ml: 0.7
 Surface Area, sq nm: 0.119

	Name	M, g	m, g	v, ml	v, ml
1	STA	284.1	0.00085	3	0.1
2	PDA	65.4	2E-5	2	0.05

Experimental Setup - Data Tab

Experiment Registry

Name	Date	User	Compound	Substance	Subphase	T	Loaded	File name
Cholesterol.mps	15.06.2005 9:57						Yes	D:\s L
DimethylEster.mps	21.06.2005 10:13						No	D:\s L

Data set: 2

Registry entry control: [Left Arrow] [Right Arrow] [Up Arrow] [Down Arrow] [Load] [Unload]

Data sheet


	EPos, mm	TA, mm2	dTA, mm2	MA, mm2	SP, mN/m	t, s	PS, mN/m
1	173.722	120.001	0.000	0.164	54.200	501.000	9.946
2	172.733	121.385	-1.354	0.165	45.900	504.900	8.704
3	171.745	122.710	-1.354	0.188	42.000	508.900	7.882
4	170.756	124.064	-1.354	0.190	38.900	512.900	7.380
5	169.768	125.418	-1.354	0.192	37.000	516.900	7.097

Graph

X Axis: Area Per Molecule (MA)
 Left Y Axis: Surface pressure (SP)
 Right Y Axis: Barrier Position (EPos)
 Plot data

Fig. 2.6 – Experimental setup panel with active tab Description (upper) and Data Sheet (lower).

- controls for loading and unloading experiment sets (group –Registry entry control–). Arrows provide switching between loaded sets in table –Experiment Registry–. Button adds new experiment set to the registry by opening file on hard disk. Inner file format for measured data storing is *.mps (MultiPurpose Set) that allows to keep unlimited number of experimental series with arbitrary data organization. Button deletes current experiment set from the list. After pressing button a confirmation will be asked for deleting file with experiment set from hard disk:



If you will leave the file on hard disk (pressing button), a confirmation for deleting the entry off the program registry will be asked:



After pressing button , the record will be deleted from the program registry (list of experiment sets).

Button / serves for loading current experiment sets in the registry if it was inactive or unloading if it was previously loaded;

- plotting of measured data as a graph in specialized window (group Graph). Selectors in the group assign any data type available in the data sheet for bottom (X Axis), left (Left Y Axis) and right (Right Y Axis) axes. To bring the graph window up and create the plot, press button at the group bottom.

2.3 Control panels

Device for film application operates under management from control panels (Fig. 2.7). They are specialized for performing following procedures:

- panel **Deposition Control** – film deposition onto substrate with feedback control and experimental data plotting (Fig. 2.7 a);
- panel **Dipping Control** – programmable dipping of a substrate into liquid without feedback control (Fig. 2.7 b);
- panel **Isotherm Control** – isotherm curve measurement and plotting without deposition (i.e. without dipping substrate into liquid) (Fig. 2.7 c).






Fig. 2.7 – Control panels of deposition (a), dipping (b), isotherm plotting (c), manual control with active tab of sensor zero point assignment (d), manual control with active tab of setting for barrier movement speed (e) and manual control with active tab of setting for dipper movement speed (f).




Switches **Rec** in panels **Deposition Control** and **Isotherm Control** when activated provide automated recording of the measured experimental data in the program registry appeared then in list of experiments in panel Experimental Setup and saving collected data in file placed in working directory.

After setting necessary operating parameters, procedure starts at pressing button **Go** in corresponding control panel. The running procedure can be manually terminated by pressing neighboring button **Stop**.

Panel **Manual Control Unit** displays in its upper part current readings of surface pressure sensor, positions and speed of barrier (group –Trough–), positions and speed of dipper (group –Dipper–), readings of temperature transducers (group –Temperature–) (Fig. 2.7 d–f).

Lower part of the panel contains buttons for manual control over the device stepper motors (group –Drives control–) (Fig. 2.7 d–f). They allow to move barrier and dipper mechanism directly without feedback control or dipping program. This function is intended for fast movement of the barrier or hang arm (with or without the attached sample) to position defined by the operator, for example for the trough cleaning or sample change/installation. Control buttons have following functions:

-  (up arrow) – start lifting of the dipper mechanism;
-  (down arrow) – start lowering of the dipper mechanism;

-  (left arrow) – start barrier movement to the left;
-  (right arrow) – start barrier movement to the right;
-  (a stop sign) – termination of the started motion.

Group –Equipment Settings– in panel **Manual Control Unit** contains three tabs: for sensor zero point assignment (tab *Balance*) (Fig. 2.7 d), barrier movement speed setting (tab *Barrier*) (Fig. 2.7 e), and dipper movement speed setting (tab *Dipper*) (Fig. 2.7 f). In tab *Dipper* additional buttons provide following functions under manual movement mode:

- **[Zero]** – assignment of zero point for dipper hang arm position. To set zero point, move manually the dipper hang arm to necessary position and press button **[Zero]** in the tab;
- **[Up]** – assignment of upper limit point for dipper hang arm position. To set upper limit point, move manually the dipper hang arm to necessary position and press button **[Up]** in the tab;
- **[Down]** – assignment of lower limit point for dipper hang arm position. To set lower limit point, move manually the dipper hang arm to necessary position and press button **[Down]** in the tab.

All positions manually assigned in panel **Manual Control Unit** will be automatically set as corresponding values in panel **Dipping Control** for next procedures.

Performing deposition, additional settings for feedback control should be made in specialized floating panel **Deposition feedback** (Fig. 2.8). Sliders in group –Feedback– are used for:

- Sensitivity – restriction of the sensor sensitivity. To avoid feedback system response to noise signals, it is better to decrease sensitivity. However, too low sensitivity will result in step-like deposition. Therefore it is necessary to find a balance between high and low feedback sensitivity levels;
- Limit velocity UP – restriction of the barrier motion speed when dipper moves upward. Barrier movement algorithm provides for dynamic speed variation defined by target distance and feedback signal magnitude. The slider allows to constrain maximum available speed of the barrier and avoid its too fast acceleration;
- Limit velocity DOWN – restriction of the barrier motion speed when dipper moves downward.




Fig. 2.8 – Control panel for adjustment of feedback parameters at deposition procedure.

Group –Deposition– of panel **Deposition feedback** contains buttons for:

- **[Zero]** – assignment of zero point for dipper hang arm position. To set zero point, move manually the dipper hang arm to necessary position and press button **[Zero]** in the panel;
- **[Up]** – assignment of upper limit point for dipper hang arm position. To set upper limit point, move manually the dipper hang arm to necessary position and press button **[Up]** in the panel;
- **[Down]** – assignment of lower limit point for dipper hang arm position. To set lower limit point, move manually the dipper hang arm to necessary position and press button **[Down]** in the panel.

All positions manually assigned in panel **Deposition feedback** will be automatically set as corresponding values in panel **Deposition Control** for next procedures.

Panel **Script Control** provides control over the device by means of scripts (Fig. 2.9). Tab *Script Control* serves for writing the device control script in its pane (Fig. 2.9 left). Script for the dipping procedure is automatically generated and is placed into the pane when dipping procedure is started from panel **Dipping Control**. If script already exists and it is in file on hard disk, it can be loaded

using 'Open file...' dialogue brought up by button **Load**. To save the script in the pane (typed or generated by program), use 'Save file...' dialogue brought up by button **Save**. Default file name extension for scripts is *.lds. The scripts are saved in text format and can be manually edited.




Fig. 2.9 – Panel Script Control.

Tab Time Graph presents timeline for the dipping process and its progress (Fig. 2.9 right).

2.4 Graph plotting window

When deposition, dipping or isotherm plotting procedures start, measured data are plotted in specialized graph windows (Fig. 2.10).




Fig. 2.10 – Example of window for plotting graph.

Isotherm graph is plotted at deposition and isotherm measurement procedures (Fig. 2.10). The curve is plotted as surface pressure (Y axis) vs. free surface area per a molecule (axis X). Surface pressure (Surface pressure, mN/m) is determined by the device as difference between absolute value of surface pressure for pure liquid medium (for water it is 72.8 mN/m) and current pressure of liquid free surface. Current pressure of liquid free surface is measured by surface pressure sensor. Free surface area per a molecule (Area per molecule, nm²) is calculated based on barrier position and quantity of substance applied on the liquid surface. Necessary data about the substance should be entered in group –Solution composition– in **Experimental Setup** panel (tab Description).

Deposition procedure also brings up graph window **Deposition** in addition to **Isotherm** plot. Graph **Deposition** plots dependencies of the transferred substance amount (left vertical axis Transfer) and ratio R/Z (transferred substance amount R to number of deposited layers Z, right vertical axis R/Z) on number of deposited layers (horizontal axis).

Dipping procedure brings up graph window **Dipping** where timeline of the dipping process is plotted. Timeline of the set dipping procedure and its progress can be also viewed in tab Time Graph of **Script Control** panel (Fig. 2.9 right).

Graph window is also brought up for plotting earlier collected data using the utility in group Graph of tab Data in **Experimental Setup** panel (Fig. 2.6 lower).

2.5 Configuration panels

Configuration of hardware and program itself for correct operation is performed using specialized panels.

Before the device operation, hardware settings should be configured for correct connection between host computer and control electronic unit. This task is performed in panel Setup (Fig. 2.11) of the program that can be brought up by menu command *Device > Device settings*.





Fig. 2.11 – Hardware configuration panel.

Settings made in the panel **Setup** relate to the serial communication port of the host computer (COM port). Commonly used values for the parameters are as below:

Port	COM1/2/3/4 (depends on the host PC configuring)
Baud rate	57600
Data bits	8
Stop bits	1
Parity	None
Flow control	None

Panel **Calibrate Device** provides access to calibration procedures for the device measuring system (Fig. 2.12). It allows correctly translate relative units used by the system to physical values of distance, area and substance amount. To bring the panel up, use menu command *View > Calibrate* or press corresponding button  in the program toolbar.




Fig. 2.12 – Panel of device calibration.

Group –Current status– shows current readings of the sensor ADC and number of steps performed by step motor. Group –Balance– contains buttons for setting *minimum* and *maximum* values of the surface pressure sensor measurement range (see also §3.2).

To set minimum value, lift the surface pressure sensor up so that sensor plate appeared in air and press button **Top level** ('top level' means that sensor plate is in upper position, i.e. in air). In pane from the button right manually enter number that will be considered as minimum value (usually it is '0').

To set maximum value, lower the surface pressure sensor on its stand so that sensor plate sinks into clean liquid and press button **Bottom level** ('bottom level' means that sensor plate is in lower

position, i.e. is sunk in liquid). In pane from the button right manually enter number that will be considered as maximum value (if liquid is distilled water, the number is '72.8').

Group –Area– provide automated calibration of the liquid surface area. After pressing button Auto calibrate area, barrier will automatically go to leftmost and then to rightmost position till limit stops to define the stepper motor step values corresponding to the biggest (leftmost) and the smallest (rightmost) areas of liquid surface. In the panes to the button right manually enter numbers corresponding to physical values of the biggest (upper pane) and the smallest (lower pane) areas of liquid surface. These values are 363 sq.cm (the biggest area) and 108 sq.cm (the smallest area).

To remember the set calibration data, press button **Store calibration** at the panel bottom. If calibration of minimum and maximum was already done with the device control electronic unit, you can transfer these data into program by pressing button **Get calibration from device**. However, area calibration can be made in the program only.

General program settings can be made in panel **Settings** brought up by menu command *File > Settings* (Fig. 2.13). The panel provides adjustment of program options for communication with the device and data accuracy (digits after decimal point) in tab **Data** and selection of the graphical interface language in tab **Interface**.




Fig. 2.13 – Panel of general program settings.

Language resources for the program interface are provided by specialized text files in subfolder **Language** located in the program folder. Used file name extension of the resource file is *.lng. New language resources can be added to the program by editing text of existing files (it is recommended to keep existing files without change and edit their copy with different name if necessary).

3 OPERATIONS WITH CONTROL ELECTRONIC UNIT

3.1 Sensor adjustment

Before the device operation, its measuring system should be properly tuned. First step in the tuning routine is mechanical adjustment of the photodetector position. To start this procedure, enter laser adjustment mode on control electronic unit by pressing button **Enter** on the keypad (before entering, make sure that the unit is under initial mode, i.e. prompt `normal mode` is in lower line of the unit display). After activation of laser adjustment mode, display will show processed output signals in upper line and prompt `<A+B> spot <A-B>` in lower line, for example:

1.31 V	0.16 V
<A+B>	spot <A-B>

Readings in upper line of the display show:

- **Left number** `<A+B>` – integral signal from the detection system photodetector.
This value is a sum of signals from upper and lower segments of the photodetector and says about general intensity of the signal. Higher value means higher intensity of the signal and says about better targeting of the laser beam onto whole sensitive area of the photodetector;
- **Right number** `<A-B>` – differential signal from the detection system photodetector.
This value is a difference between signals from upper and lower segments of the photodetector and says about targeting of the reflected laser beam spot to the photodetector center. The best targeting of the reflected laser beam spot gives value '0' that says about correct spot position right on the photodetector center line.

Second pressing of button **Enter** on the keypad when system is under laser adjustment mode toggles output signals format to ADC values from lower and upper segments of the photodetector. Lower line will change to `A photo diode B` to prompt about signals from lower `A` or upper `B` segment of the photodetector, for example:

1601	725
A photo diode B	

This mode shows distribution of cumulative output signal measured on lower A and upper B segments of the photodetector. The best targeting of the reflected laser beam spot gives equal values for segment output signals (i.e. `<A>=`) that says about correct spot position right on the photodetector center line.

Mechanical adjustment of the photodetector position provides correct targeting of the photodetector sensitive area to laser beam reflected from the flexible cantilever rear side (Fig. 1.3). To adjust the photodetector position, using a flat screwdriver, turn the adjustment screw in a hole on upper side of surface pressure sensor (Fig. 1.3). Correct adjustment results in positioning of the photodetector sensitive area center line right opposite center of spot of laser beam reflected from the flexible cantilever rear side.

This procedure should be monitored on the control electronic unit display (under laser adjustment mode). If processed output signals are shown in the display, the best adjustment of the photodetector position is reached when integral signal (left number) is as big as possible and differential signal (right number) is '0'. If the display shows segment signal values, the best adjustment of the photodetector position is reached when left number (A, lower segment signal) equals right number (B, upper segment signal). Note that the values should be as high as possible.

During primary adjustment of the laser detection system (i.e. photodetector position), it is recommended to check visually position of laser beam spot on the photodetector sensitive area. It should be on right group of the light-sensitive segments and targeted on horizontal separating line as shown in Fig. 3.1. To maximize output signal, the spot should not overlap vertical separating line.




Fig. 3.1 – Correct positioning of laser beam spot on the photodetector sensitive area.

After finishing adjustment of the photodetector position it is recommended to check range of the sensor output signal. This task is fulfilled with unloaded sensor cantilever by slight turning the photodetector adjustment knob clockwise and counterclockwise from its adjusted state. Readings on the display (under laser calibration mode) should change when screw is turned in both directions. After the check, set the photodetector adjusted position again (i.e. till equal signals from the segments or zero differential signal).

To leave the laser adjustment mode of measuring system, press button **Esc** on the keypad. Initial display mode will return.

3.2 Sensor calibration

Calibration of surface pressure sensor is second step of the device measuring system tuning. It should be done before operating the device. Aim of this procedure is calibration of the sensor output signal range to provide the measured data reliability. It is recommended to calibrate sensor before each experiment and every time after the Wilghelmy plate change and change of liquid in trough. Sensor calibration procedure includes two stages: (1) calibration of minimum value of the surface pressure sensor measurement range and (2) calibration of maximum value of measurement range of surface pressure sensor.

To enter calibration mode to define minimum value, being in initial mode, press button **1** on the keypad. For example:

0.0 0
calibration min

To enter calibration mode to define maximum value, being in initial mode, press button **2** on the keypad. For example:

72.8	50
calibration max	

Switch between these two modes can be performed by pressing corresponding button on the keypad without return to initial mode.

Readings in upper line on display show:

- <left number> – minimum/maximum value of the surface pressure in physical units (mN/m);
- <right number> – relative deflection of the spot from the central position in per cents.

First stage – minimum value calibration. To fulfill this stage, activate mode of minimum value calibration (being in initial mode, press button **1** on the keypad) and lift the surface pressure sensor up so that sensor plate appears in air. The prompt on the display will be like this:

0.0	0
calibration min	

To edit number for physical minimum value on the display (usually it is '0'), move the cursor in upper line with buttons **←** and **→** on the keypad to position of edited digit and using up arrow **↑** and down arrow **↓** buttons on the keypad increase or reduce the value. For example:

0.0	0
editing minimum	

After reaching necessary value of the parameter, press button **Enter** on the keypad to enter it into the device memory. In lower line of display prompt **data are stored** will be shown:

0.0	0
data are stored	

Second stage – maximum value calibration. To calibrate maximum value, activate mode of maximum value calibration (being in initial mode or in mode of minimum value calibration, press button **2** on the keypad) and lower the surface pressure sensor on its stand so that sensor plate sinks into clean liquid (in the trough or other vessel). The prompt on the display will be like this:

72.8	50
calibration max	


If left number is different from the value typical for used liquid (for clean distilled water surface pressure is 72.8 mN/m), edit this number manually on the display: move the cursor in upper line with buttons **←** and **→** on the keypad to position of edited digit and using up arrow **↑** and down arrow **↓** buttons on the keypad increase or reduce the value. For example:

72.4	50
editing maximum	

After reaching necessary value of the parameter, press button **Enter** on the keypad to save it into the device memory. In lower line of display prompt **data are stored** will be shown:

72.8	50
data are stored	



After finishing both stages of the calibration procedure with control electronic unit it is recommended to transfer the adjusted parameters to control software. For that, bring up panel



Calibrate Device using menu command *View > Calibrate* or button  in the program toolbar and then press button Get calibration from device at the panel bottom. All necessary data will be then updated in the program automatically.



To leave the sensor calibration mode of measuring system, press button **Esc** on the keypad. Initial display mode will return.



3.3 Manual control of motors

To move barrier or dipper hang arm into arbitrary position, manual control of the stepper motors available with the control electronic unit can be employed. For this purpose arrow buttons on the keypad should be used. Directions of the arrows on the keypad correspond to directions of movements fulfilled at their pressing.

For manual barrier movement, left arrow  and right arrow  buttons on the keypad are used:

- left arrow button  starts barrier movement to left side (free area decrease). The motion is performed till barrier reaches leftmost position and can be manually terminated by button **Esc** on the keypad;
- right arrow button  starts barrier movement to right side (free area increase). The motion is performed till barrier reaches rightmost position and can be manually terminated by button **Esc** on the keypad.

For manual movement of the dipper hang arm, up arrow  and down arrow  buttons on the keypad are used:

- up arrow button  runs dipper hang arm lifting. The motion is performed only while the button is pressed and stops when the button is released;
- down arrow button  runs dipper hang arm lowering. The motion is performed only while the button is pressed and stops when the button is released.

When motors are manually controlled, display shows number of the executed steps (in upper line) and corresponding prompt about the movement direction (in lower line), for example:

1804	-1178
motor goto right	

At the motor movement stop, prompt in lower line changes to motor stopped, for example:

1682	-1196
motor stopped	

To display information about current position of the device stepper motors (in performed steps), press button **5** on the keypad.

4 PRACTICAL GUIDE TO OPERATION WITH EXPERIMENTAL DEVICE FOR FILM APPLICATION LT-103

Experimental device for film application LT-103 is intended for studies of surface active properties of amphiphilic compounds, formation of monomolecular films on aqueous surface and modification of solid substrates by mono- and multimolecular Langmuir–Blodgett films.

At the operation with LT-103 device, following stages usually should be fulfilled:

1. Thoroughly (with accuracy not worse than $0.1 \cdot 10^{-4}$ g) weigh the studied substances and prepare their 1–0.1 mMole solutions in appropriate solvent (e.g. chloroform, benzole, hexane).
2. Switch the control electronic unit on, run control software on host PC, and thoroughly calibrate the surface pressure sensor (it operates based on measurement of surface tension by Wilghelmy technique).
3. Apply the studied compound (its solution) on to the aqueous surface.
4. Get an isotherm in coordinates 'surface pressure – area per molecule, π -A'.
5. Choose the deposition method and construct the film of intended thickness on he substrate.

Design of LT-103 device also allows depositing compounds from solution by cyclic dipping of the substrate into the liquid at preset speed according to the assigned script. Below, you can find details for each of these steps of the device operation.

4.1 Switching the device on and calibration

Surface pressure sensor has sensitivity of order 0.02 mN/m and it is recommended to calibrate it before every isotherm measurement.



ATTENTION! Do not expose photosensitive element of the sensor to direct sunlight to avoid drifting of zero point of the surface pressure signal.

The sensor should be calibrated in two positions. Fully moist Wilghelmy plate being completely in the air (out of contact with liquid) corresponds to zero surface tension.

1. After filling the trough with liquid (usually with water), dip the surface pressure unit so that the Wilghelmy plate hung on the sensitive cantilever immersed in the liquid in got fully moist (e.g. for 3–5 min).
2. Lift the sensor so that the plate appeared completely in the air.
3. Press button **UP** in 'Calibrate device' window (Fig. 4.1) to store the value of the moist plate weight and assign zero level to this value.
4. Lower the sensor unit to sink the plate with its lower edge into water.
5. Press button **DOWN** in 'Calibrate device' window to store the surface tension value corresponding to clean aqueous surface, value 72.8 will be assigned to current ADC reading.




Fig. 4.1 – Main windows of the LT-103 control software.

In panel 'Manual Control Unit' (Fig. 4.2), parameter Surface Tension will get value 72.8 mN/m, and parameter SP (Surface Pressure, Fig. 4.1) in General Indicator panel will be 0 according to the equation

$$\pi = -(\sigma - \sigma_0) = 0,$$

where σ_0 is surface tension of clean water surface 72.8 mN/m at 25 °C; σ is current value of surface tension during the monolayer compression.

After pressing button **Store Calibration** in window 'Calibrate Device', Fig. 4.1, the program transfers the assigned parameters to the control electronic unit to store them there.

There is an option to calibrate the device based on the area using the standard substance with known value of minimal area per a molecule in close-packed monolayer, e.g. stearic acid. For that, record initially an isotherm and determine minimum area $A = 0.205 \text{ nm}^2$.

Note that displayed value of a parameter may differ from its actual one and it can be conditioned by two reasons: wrong weighing and application of the substance onto aqueous surface (as usual) or incorrect calibration of the device. To change minimum and maximum position of the barrier, press button **Auto Calibrate Area** and following prompts in the program enter minimum and maximum barrier positions so that the trough working area was as closer to its actual size as possible.

In the device sensor, there is also an option to change the sensitive cantilever spring constant to adjust its sensitivity. After change of the cantilever free end length by moving the sliding slat on sensor unit bottom side please don't forget to readjust the laser detection system so that working stroke of the reflected spot on the photodetector during measurements appeared within photosensitive area of the latter.

Fig. 4.2 – Panel “Manual Control unit”

4.2 Measurement of compression isotherms

Before application of the working solution onto aqueous surface, manually clean thoroughly the working area. For that, pressing buttons **Left** and **Right** in ‘Manual Control Unit’ panel (Fig. 4.2) move the barrier to position of minimum working area and clean the constrained aqueous surface with a water jet pump. Then move the barrier to position of maximum working area and calibrate the surface tension sensor according to §4.1.

After application of necessary amount of substance onto aqueous surface, fill information in section Data in panel ‘Experimental Setup’ (Fig. 4.1): operator (user) name, the substance name (abbreviated or contracted), type (nature) of the substrate, composition of the subphase, temperature. Enter also number of compounds in the solution, volume of solvent, molecular weight of substance, weight and volume of sample applied onto aqueous surface. If the table in pane ‘Surface Area’ is filled in correctly, maximum area per a molecule for given experiment will be automatically calculated and displayed.

In case of multicomponent solutions, operator instead of filling in value ‘Volume of Solvent’ should enter in the table the data describing solutions of individual compounds BEFORE their mixture, namely molecular weights (M, g), sample weights (m, g) and solvent volumes (V, ml) used for the sample preparation. In the last column, enter the aliquot volumes taken for the component mixture and preparation of working solution with necessary proportion of the components. The program will automatically add these values and calculate the total volume of working solution in field ‘Volume of Solvent’.

Next stage of the experient is obtaining of isotherm under parameters set in panel ‘Isotherm Control’ (Fig. 4.3) that can be activated by selecting corresponding tab in panel ‘Manual Control Unit’.

The operator can assign necessary number of cycles of the isotherm recording (usually 1), enable or disable the isotherm recording at the barrier backward motion (option 'Backward'), enable automatic recording of the isotherm by switch 'Rec' (red color corresponds to enabled option). Other Recording Options allow to specify how to measure the isotherm: till reaching the preset value target parameter, till manual stop, from preset value of target parameter till manual stop, from preset value of target parameter during the assigned time interval, under completely manual control by operator (using buttons **Left** and **Right** in 'Manual Control Unit' panel). The target parameter can be chosen between surface pressure, the trough surface area and barrier position. The operator also assigns compression type (Constant or with Delay), compression speed and the barrier step defining points where area per molecule and corresponding surface pressure will be measured. The less value 'Barrier Step', the more number of measurement points recorded in data file and used for diagram plotting. Automatic measurement of isotherm begins at pressing button **Go**.




Fig. 4.3 – The program windows used during investigation of substance monolayer behavior under compression.

4.3 Deposition of mono- and multimolecular films

To deposit quality mono- and multimolecular films, the operator should use appropriate surfactants and choose optimal deposition pressure value.

Preparatory tasks before the deposition start include cleaning of the aqueous surface from contaminations, the sensor calibration and application of a substance onto aqueous surface. Deposition process parameters should be set in panel 'Deposition Control' activated by selecting corresponding tab in panel 'Manual Control Unit' (Fig. 4.4).




Fig. 4.4 – The program windows and panels used for control over the film deposition process.

Available are three deposition modes: automatic by traditional vertical (LB) method of formation of structures of X-, Y-, Z-types, horizontal precipitation of monolayer and film deposition under manual control. In case of vertical deposition method and formation of multimolecular films, assigned should be number of cycles, entered the substrate surface area (both sides), deposition type, way of deposition (with constant speed or by steps with delay) and other parameters like barrier speed, pressure of deposition, speeds of the substrate movement up and down, regimes of drying/moistening. additionally can be set step of the substrate motion and delay at step-like dipping/lifting of the substrate.



Depending on deposition speed, construction of films of three types is available: X- and Z-types of non-centrosymmetrical LB films and Y-type of centrosymmetrical LB films (Fig. 4.5). To deposit X-type film, substrate slowly dips through the monomolecular film at certain pressure and then quickly moves up. To deposit Z-type film, all is vice versa – substrate quickly dips through monomolecular film at certain pressure and then slowly moves up. Bilayered centrosymmetrical LB films can be constructed at various speeds of substrate motion through free surface of liquid in trough.




Fig. 4.5 – Schematics of Langmuir-Blodgett film of different structure.

Activated option 'Isotherm Control' means that during the deposition process not only the process progress is displayed (diagrams showing the film transfer degree) but also surface pressure changes during the deposition process will be plotted (Fig. 4.6).

From Fig. 4.6 it is seen that in case of deposition of behenic acid, quantitative transfer of the monolayer onto substrate takes place with transfer ratio 1. Right lower image in Fig. 4.6 displays feedback control panel for system 'substrate-barrier'.

Selection of HP mode in section 'Deposition Type' will result in compression of monolayer by barrier till the set value of surface pressure. Then the operator should drain the water to lower the free surface level below the substrate placed on a specialized rig (table) in the dipping well.




Fig. 4.6 – Example of transfer process visualization for behenic acid film.

4.4 Data processing

Operator can save and restore data using tab 'Description' in window 'Experimental Setup' (Fig. 4.7) or open recorded data files via menu 'File > Open'.




Fig. 4.7 – Main window displaying recorded isotherms and used for data file searching and opening.

In the latter case, pane Data Sheet displays the recorded data. Operator can choose mode (Isotherm or Deposition) in field 'Data Set', assign axes (X and Y) and plot the diagram by pressing button **Plot Data**. After that the diagram will be displayed. The diagram properties can be adjusted in additional panel 'Graph Properties' brought out at double click in any place of the diagram window (Fig. 4.8).

Tab 'Graph' allows to change the diagram title, its color and font. Additionally can be changed the diagram background color, axes color, displayed or hidden grid (option 'Grid'), adjusted the grid color, displayed or hidden the diagram legend (option 'Legend') (Fig. 4.9).

There is also an option to display multiple graph lines in the same diagram area (Fig. 4.10) that enables direct data comparison. To add additional lines into the diagram, use section 'Lines' in tab 'Axis' > 'Left' or 'Axis' > 'Right'. Operator can select name and assign data string to the selected line (correspondingly, selectors 'Experiment' and 'Data' in section 'Source'). Section 'Line Properties' allows to adjust the selected line displaying parameters.




Fig. 4.8 – Example of editing the diagram window.




Fig. 4.9 – Example of diagram with legend (list of axes types).




Fig. 4.10 – Example of plotting multiple diagram in one field.



MICROTESTMACHINES Co.

2014